xylem
Let's Solve Water

Xylem operates under the brands of YSI, SonTek, Aanderaa, WaterLOG, and MJK

YSI Incorporated
1700/1725 Brannum Lane
Yellow Springs, OH 45387
USA
tel: 937-767-7241
fax: 937-767-9320

# Packing List



| Delivery Number | Packing Slip Number | Date Shipped |
|---|---|---|
| 1393648 | YSI-388164 | 09-MAR-15 |
| Shipped Via: FEDEX-Air-FedEx 2nd Day | | |

## CERTIFICATE OF COMPLIANCE

***This is to certify that the materials, processes and finished products were controlled, tested, and accepted in accordance with our quality system and the applicable specifications.***

**Ron Geis**
**Vice President & General Manager**

**SOLD TO**

NATIONAL PARK SERVICE
3100 NATIONAL PARK RD
JUNEAU, AK 99801, United States

**SHIP TO**

NATIONAL PARK SERVICE
3100 NATIONAL PARK RD
JUNEAU, AK 99801, United States

| PICKUP DATE | PAYMENT TERMS | FREIGHT TERMS | SR NUMBER | F.O.B | TAX NAME | TAX NO. | Supplier No. |
|---|---|---|---|---|---|---|---|
| 09-MAR-15 | Credit Card | Prepaid | 310467 | Shipping Point | Exempt | | |

| ORDER LINE NO. | YSI ITEM NO. | ORDER NO. | CUSTOMER ITEM NO. | PO NO. | PO Line# | DESCRIPTION | HS NO./ Country Of Origin | UOM | QTY ORDERED | QTY SHIPPED | BACK ORDERED |
|---|---|---|---|---|---|---|---|---|---|---|---|
| 10.1 | R6136 | 657826 | | CREDITCARD | | Repair, Model 6136<br><br>6136 TURBIDITY PROBE SN# 10C101562<br><br>REPORTED SYMPTOM:<br>PREVENTATIVE MAINTENANCE<br><br>TESTED AND CALIBRATED PROBE.<br>ALL UPDATES IN PLACE<br>REPLACED QUAD SEAL.<br>REPLACED O-RINGS.<br>TEST WIPER FUNCTION, NO DEFECT FOUND.<br>PROBE READS: 126.8 IN 126 STANDARD. | 9801.10.0000 | EA | 1 | 1 | 0 |

-*For terms and conditions refer to www.ysi.com/terms_conditions*
-These commodities, technology or software were exported from the United States in accordance with the Export Administration Regulations. Diversion contrary to U.S. law is prohibited.

xylem
Let's Solve Water

Xylem operates under the brands of YSI, SonTek, Aanderaa, WaterLOG, and MJK

YSI Incorporated
1700/1725 Brannum Lane
Yellow Springs, OH 45387
USA
tel: 937-767-7241
fax: 937-767-9320

# Packing List



| Delivery Number | Packing Slip Number | Date Shipped |
|---|---|---|
| 1393648 | YSI-388164 | 09-MAR-15 |
| Shipped Via: FEDEX-Air-FedEx 2nd Day | | |

| ORDER LINE NO. | YSI ITEM NO. | ORDER NO. | CUSTOMER ITEM NO. | PO NO. | PO Line# | DESCRIPTION | HS NO./ Country Of Origin | UOM | QTY ORDERED | QTY SHIPPED | BACK ORDERED |
|---|---|---|---|---|---|---|---|---|---|---|---|
| | | | | | | FINAL TEST, MEETS YSI SPECIFICATIONS. | | | | | |
| 12.1 | R6920V2 | 657826 | | CREDITCARD | | Repair,Model 6920V2<br><br>MODEL# 6920 SN# 08H100353<br><br>REPORTED SYMPTOM: P.M.<br><br>TEST AND CALIBRATE BOARD, NO DEFECT FOUND.<br>TEST RTC BATTERY, NO DEFECT FOUND.<br>FIRMWARE UP TO DATE AT 3.10.<br>PERFORM CURRENT DRAW TEST. PASSED SLEEP, IDLE AND RUN MODE TESTING.<br>TESTED WITH BATTERIES AND POWER SUPPLY.<br><br>TEST BULKHEAD FOR STRIPPED PORTS AND HIGH IMPEDANCE LEAKAGE. NO DEFECT FOUND.<br>TEST BCR CONNECTOR AND BATTERY COMPARTMENT. NO DEFECT FOUND.<br><br>TEST CONDUCTIVITY/TEMPERATURE PROBE(SN#11H100545) TEMPERATURE READS 20.55°C IN A 20.5°C BATH WITH A CELL CONSTANT OF 4.84988 PROBE READS ZERO SPCOND. IN AIR.<br><br>TEST PH PROBE(SN#13A) REPLACE WITH NEW DUE TO OUT OF SPEC READINGS.<br>REPLACED AND TESTED PH PROBE(SN#15A) NO DEFECT | 9801.10.0000<br>United States | EA | 1 | 1 | 0 |

*-For terms and conditions refer to www.ysi.com/terms_conditions*

-These commodities, technology or software were exported from the United States in accordance with the Export Administration Regulations. Diversion contrary to U.S. law is prohibited.

xylem
Let's Solve Water

Xylem operates under the brands of YSI, SonTek, Aanderaa, WaterLOG, and MJK

YSI Incorporated
1700/1725 Brannum Lane
Yellow Springs, OH 45387
USA
tel: 937-767-7241
fax: 937-767-9320

# Packing List



| Delivery Number | Packing Slip Number | Date Shipped |
|---|---|---|
| 1393648 | YSI-388164 | 09-MAR-15 |
| Shipped Via: FEDEX-Air-FedEx 2nd Day | | |

| ORDER LINE NO. | YSI ITEM NO. | ORDER NO. | CUSTOMER ITEM NO. | PO NO. | PO Line# | DESCRIPTION | HS NO./ Country Of Origin | UOM | QTY ORDERED | QTY SHIPPED | BACK ORDERED |
|---|---|---|---|---|---|---|---|---|---|---|---|
| | | | | | | FOUND. READINGS BELOW:<br>7: -0.2mV<br>10: -175.9mV<br>4: 167.5mV SLOPE:<br>175.7mV<br><br>ALL PORTS TEST= FUNCTIONAL.<br><br>REPLACED O-RINGS.<br><br>FINAL TEST, MEETS YSI SPECIFICATIONS. | | | | | |
| 14.1 | R6150 | 657826 | | CREDITCARD | | Repair, Model 6150<br><br>6150 ROX ODO PROBE SN# 08H101103<br><br>REPORTED SYMPTOM:<br>PREVENTATIVE MAINTENANCE<br><br>TESTED AND CALIBRATED PROBE.<br>ALL UPDATES IN PLACE<br>REPLACED QUAD SEAL.<br>REPLACED O-RINGS.<br>REPLACED MEMBRANE WITH PART#606155,LOT#15B100357<br>TEST WIPER FUNCTION, NO DEFECT FOUND.<br>PROBE READS: 100% IN AERATED BATH.<br>GAIN: 1.07739<br>FINAL TEST, MEETS YSI SPECIFICATIONS. | 9801.10.0000 | EA | 1 | 1 | 0 |
| 2.1 | R6920V2 | 657826 | | CREDITCARD | | Repair,Model 6920V2<br><br>MODEL# 6920V2 SN# 08H100354 | 9801.10.0000<br>United States | EA | 1 | 1 | 0 |

*-For terms and conditions refer to www.ysi.com/terms_conditions*

-These commodities, technology or software were exported from the United States in accordance with the Export Administration Regulations. Diversion contrary to U.S. law is prohibited.

xylem
Let's Solve Water

Xylem operates under the brands of YSI, SonTek, Aanderaa, WaterLOG, and MJK

YSI Incorporated
1700/1725 Brannum Lane
Yellow Springs, OH 45387
USA
tel: 937-767-7241
fax: 937-767-9320

# Packing List



| Delivery Number | Packing Slip Number | Date Shipped |
|---|---|---|
| 1393648 | YSI-388164 | 09-MAR-15 |
| Shipped Via: FEDEX-Air-FedEx 2nd Day | | |

| ORDER LINE NO. | YSI ITEM NO. | ORDER NO. | CUSTOMER ITEM NO. | PO NO. | PO Line# | DESCRIPTION | HS NO./ Country Of Origin | UOM | QTY ORDERED | QTY SHIPPED | BACK ORDERED |
|---|---|---|---|---|---|---|---|---|---|---|---|
| | | | | | | REPORTED SYMPTOM: P.M.<br><br>TEST AND CALIBRATE BOARD, NO DEFECT FOUND.<br>TEST RTC BATTERY, NO DEFECT FOUND.<br>FIRMWARE UP TO DATE AT 3.10.<br>PERFORM CURRENT DRAW TEST. PASSED SLEEP, IDLE AND RUN MODE TESTING.<br>TESTED WITH BATTERIES AND POWER SUPPLY.<br><br>TEST BULKHEAD FOR STRIPPED PORTS AND HIGH IMPEDANCE LEAKAGE. NO DEFECT FOUND.<br>TEST BCR CONNECTOR AND BATTERY COMPARTMENT. NO DEFECT FOUND.<br><br>TEST CONDUCTIVITY/TEMPERATURE PROBE(SN#13A100461)<br>TEMPERATURE READS 19.83°C IN A 19.8°C BATH WITH A CELL CONSTANT OF 4.91845 PROBE READS ZERO SPCOND. IN AIR.<br>TEST PH PROBE(SN#14F) NO DEFECT FOUND. READINGS BELOW:<br>7: -16.9mV<br>10: -194.9mV<br>4: 153.5mV SLOPE: 178mV<br><br>ALL PORTS TEST=FUNCTIONAL<br><br>REPLACED O-RINGS.<br><br>FINAL TEST, MEETS YSI | | | | | |

*-For terms and conditions refer to www.ysi.com/terms_conditions*

-These commodities, technology or software were exported from the United States in accordance with the Export Administration Regulations. Diversion contrary to U.S. law is prohibited.

xylem
Let's Solve Water

Xylem operates under the brands of YSI, SonTek, Aanderaa, WaterLOG, and MJK

YSI Incorporated
1700/1725 Brannum Lane
Yellow Springs, OH 45387
USA
tel: 937-767-7241
fax: 937-767-9320

# Packing List



| Delivery Number | Packing Slip Number | Date Shipped |
|---|---|---|
| 1393648 | YSI-388164 | 09-MAR-15 |
| Shipped Via: FEDEX-Air-FedEx 2nd Day | | |

| ORDER LINE NO. | YSI ITEM NO. | ORDER NO. | CUSTOMER ITEM NO. | PO NO. | PO Line# | DESCRIPTION | HS NO./ Country Of Origin | UOM | QTY ORDERED | QTY SHIPPED | BACK ORDERED |
|---|---|---|---|---|---|---|---|---|---|---|---|
| | | | | | | SPECIFICATIONS. | | | | | |
| 4.1 | R6150 | 657826 | | CREDITCARD | | Repair, Model 6150<br><br>6150 ROX ODO PROBE SN# 11E101704<br><br>REPORTED SYMPTOM: PREVENTATIVE MAINTENANCE<br><br>TESTED AND CALIBRATED PROBE.<br>ALL UPDATES IN PLACE<br>REPLACED QUAD SEAL.<br>REPLACED O-RINGS.<br>REPLACED MEMBRANE WITH PART#606155,LOT#15B100357<br>TEST WIPER FUNCTION, NO DEFECT FOUND.<br>PROBE READS: 100% IN AERATED BATH.<br>GAIN: 1.10519<br>FINAL TEST, MEETS YSI SPECIFICATIONS. | 9801.10.0000 | EA | 1 | 1 | 0 |
| 6.1 | R6920 | 657826 | | CREDITCARD | | Repair,Model 6920<br><br>MODEL# 6920 SN# 08H100355<br><br>REPORTED SYMPTOM: P.M.<br><br>TEST AND CALIBRATE BOARD, NO DEFECT FOUND.<br>TEST RTC BATTERY, NO DEFECT FOUND.<br>FIRMWARE UP TO DATE AT 3.10.<br>PERFORM CURRENT DRAW TEST. PASSED SLEEP, IDLE AND RUN MODE TESTING.<br>TESTED WITH BATTERIES AND POWER SUPPLY. | 9801.10.0000 | EA | 1 | 1 | 0 |

*-For terms and conditions refer to www.ysi.com/terms_conditions*

-These commodities, technology or software were exported from the United States in accordance with the Export Administration Regulations. Diversion contrary to U.S. law is prohibited.

xylem
Let's Solve Water

Xylem operates under the brands of YSI, SonTek, Aanderaa, WaterLOG, and MJK

YSI Incorporated
1700/1725 Brannum Lane
Yellow Springs, OH 45387
USA
tel: 937-767-7241
fax: 937-767-9320

# Packing List



| Delivery Number | Packing Slip Number | Date Shipped |
|---|---|---|
| 1393648 | YSI-388164 | 09-MAR-15 |
| Shipped Via: FEDEX-Air-FedEx 2nd Day | | |

| ORDER LINE NO. | YSI ITEM NO. | ORDER NO. | CUSTOMER ITEM NO. | PO NO. | PO Line# | DESCRIPTION | HS NO./ Country Of Origin | UOM | QTY ORDERED | QTY SHIPPED | BACK ORDERED |
|---|---|---|---|---|---|---|---|---|---|---|---|
| | | | | | | TEST BULKHEAD FOR STRIPPED PORTS AND HIGH IMPEDANCE LEAKAGE. NO DEFECT FOUND. TEST BCR CONNECTOR AND BATTERY COMPARTMENT. NO DEFECT FOUND.<br><br>TEST CONDUCTIVITY/TEMPERATURE PROBE(SN#10F100521) TEMPERATURE READS 20.44℃ IN A 20.4℃ BATH WITH A CELL CONSTANT OF 4.8361 PROBE READS ZERO SPCOND. IN AIR.<br><br>TEST PH PROBE(SN#13A) REPLACE WITH NEW DUE TO OUT OF SPEC READINGS. REPLACED AND TESTED PH PROBE(SN#15A) NO DEFECT FOUND. READINGS BELOW:<br>7: -2.7mV<br>10: -178.2mV<br>4: 168.3mV SLOPE: 175.5mV<br><br>ALL PORTS TEST=FUNCTIONAL.<br><br>REPLACED O-RINGS.<br><br>FINAL TEST, MEETS YSI SPECIFICATIONS. | | | | | |
| 8.1 | R6150 | 657826 | | CREDITCARD | | Repair, Model 6150<br><br>6150 ROX ODO PROBE SN# 08H101102<br><br>REPORTED SYMPTOM: PREVENTATIVE MAINTENANCE | 9801.10.0000 | EA | 1 | 1 | 0 |

*-For terms and conditions refer to www.ysi.com/terms_conditions*

-These commodities, technology or software were exported from the United States in accordance with the Export Administration Regulations. Diversion contrary to U.S. law is prohibited.

xylem
Let's Solve Water

Xylem operates under the brands of YSI, SonTek, Aanderaa, WaterLOG, and MJK

YSI Incorporated
1700/1725 Brannum Lane
Yellow Springs, OH 45387
USA
tel: 937-767-7241
fax: 937-767-9320

# Packing List



| Delivery Number | Packing Slip Number | Date Shipped |
|---|---|---|
| 1393648 | YSI-388164 | 09-MAR-15 |
| Shipped Via: FEDEX-Air-FedEx 2nd Day | | |

| ORDER LINE NO. | YSI ITEM NO. | ORDER NO. | CUSTOMER ITEM NO. | PO NO. | PO Line# | DESCRIPTION | HS NO./ Country Of Origin | UOM | QTY ORDERED | QTY SHIPPED | BACK ORDERED |
|---|---|---|---|---|---|---|---|---|---|---|---|
| | | | | | | TESTED AND CALIBRATED PROBE.<br>ALL UPDATES IN PLACE<br>REPLACED QUAD SEAL.<br>REPLACED O-RINGS.<br>REPLACED MEMBRANE WITH PART#606155,LOT#15B100357<br>TEST WIPER FUNCTION, NO DEFECT FOUND.<br>PROBE READS: 100% IN AERATED BATH.<br>GAIN: 1.12222<br>FINAL TEST, MEETS YSI SPECIFICATIONS. | | | | | |

**GROSS WEIGHT(LB) :** 18.20 **NET WEIGHT (LB) :** 18.20 **VOLUME(CC) :** 0 **TOTAL CARTONS:** 0

**ADDITIONAL INFO:** 615558369683

**ADDITIONAL NOTE:** **Shipping Note: Invoice Note:**

Tracking Number: 615558369683

*-For terms and conditions refer to www.ysi.com/terms_conditions*

-These commodities, technology or software were exported from the United States in accordance with the Export Administration Regulations. Diversion contrary to U.S. law is prohibited.