



Xylem operates under the brands of YSI, SonTek, Aanderaa, WaterLOG, and MJK

YSI Incorporated
1700/1725 Brannum Lane
Yellow Springs, OH 45387
USA
tel: 937-767-7241
fax: 937-767-9320

Packing List

Page 1 of 7

Delivery Number	Packing Slip Number	Date Shipped
1393648	YSI-388164	09-MAR-15
Shipped Via: FEDEX-Air-FedEx 2nd Day		

CERTIFICATE OF COMPLIANCE

This is to certify that the materials, processes and finished products were controlled, tested, and accepted in accordance with our quality system and the applicable specifications.

Ron Geis
Vice President & General Manager

SOLD TO

NATIONAL PARK SERVICE
3100 NATIONAL PARK RD
JUNEAU, AK 99801, United States

SHIP TO

NATIONAL PARK SERVICE
3100 NATIONAL PARK RD
JUNEAU, AK 99801, United States

PICKUP DATE	PAYMENT TERMS	FREIGHT TERMS	SR NUMBER	F.O.B	TAX NAME	TAX NO.	Supplier No.
09-MAR-15	Credit Card	Prepaid	310467	Shipping Point	Exempt		

ORDER LINE NO.	YSI ITEM NO.	ORDER NO.	CUSTOMER ITEM NO.	PO NO.	PO Line#	DESCRIPTION	HS NO./ Country Of Origin	UOM	QTY ORDERED	QTY SHIPPED	BACK ORDERED
10.1	R6136	657826		CREDITC ARD		Repair, Model 6136 6136 TURBIDITY PROBE SN# 10C101562 REPORTED SYMPTOM: PREVENTATIVE MAINTENANCE TESTED AND CALIBRATED PROBE. ALL UPDATES IN PLACE REPLACED QUAD SEAL. REPLACED O-RINGS. TEST WIPER FUNCTION, NO DEFECT FOUND. PROBE READS: 126.8 IN 126 STANDARD.	9801.10.0000	EA	1	1	0

-For terms and conditions refer to www.ysi.com/terms_conditions

-These commodities, technology or software were exported from the United States in accordance with the Export Administration Regulations. Diversion contrary to U.S. law is prohibited.





Xylem operates under the brands of YSI, SonTek, Aanderaa, WaterLOG, and MJK

YSI Incorporated
1700/1725 Brannum Lane
Yellow Springs, OH 45387
USA
tel: 937-767-7241
fax: 937-767-9320

Packing List

Delivery Number	Packing Slip Number	Date Shipped
1393648	YSI-388164	09-MAR-15
Shipped Via: FEDEX-Air-FedEx 2nd Day		

ORDER LINE NO.	YSI ITEM NO.	ORDER NO.	CUSTOMER ITEM NO.	PO NO.	PO Line#	DESCRIPTION	HS NO./ Country Of Origin	UOM	QTY ORDERED	QTY SHIPPED	BACK ORDERED
12.1	R6920V2	657826		CREDITC ARD		FINAL TEST, MEETS YSI SPECIFICATIONS. Repair, Model 6920V2 MODEL# 6920 SN# 08H100353 REPORTED SYMPTOM: P.M. TEST AND CALIBRATE BOARD, NO DEFECT FOUND. TEST RTC BATTERY, NO DEFECT FOUND. FIRMWARE UP TO DATE AT 3.10. PERFORM CURRENT DRAW TEST. PASSED SLEEP, IDLE AND RUN MODE TESTING. TESTED WITH BATTERIES AND POWER SUPPLY. TEST BULKHEAD FOR STRIPPED PORTS AND HIGH IMPEDANCE LEAKAGE. NO DEFECT FOUND. TEST BCR CONNECTOR AND BATTERY COMPARTMENT. NO DEFECT FOUND. TEST CONDUCTIVITY/TEMPERATURE PROBE(SN#11H100545) TEMPERATURE READS 20.55°C IN A 20.5°C BATH WITH A CELL CONSTANT OF 4.84988 PROBE READS ZERO SPCOND. IN AIR. TEST PH PROBE(SN#13A) REPLACE WITH NEW DUE TO OUT OF SPEC READINGS. REPLACED AND TESTED PH PROBE(SN#15A) NO DEFECT	9801.10.0000 United States	EA	1	1	0

-For terms and conditions refer to www.ysi.com/terms_conditions

-These commodities, technology or software were exported from the United States in accordance with the Export Administration Regulations. Diversion contrary to U.S. law is prohibited.





Xylem operates under the brands of YSI, SonTek, Aanderaa, WaterLOG, and MJK

YSI Incorporated
1700/1725 Brannum Lane
Yellow Springs, OH 45387
USA
tel: 937-767-7241
fax: 937-767-9320

Packing List

Delivery Number	Packing Slip Number	Date Shipped
1393648	YSI-388164	09-MAR-15
Shipped Via: FEDEX-Air-FedEx 2nd Day		

ORDER LINE NO.	YSI ITEM NO.	ORDER NO.	CUSTOMER ITEM NO.	PO NO.	PO Line#	DESCRIPTION	HS NO./ Country Of Origin	UOM	QTY ORDERED	QTY SHIPPED	BACK ORDERED
						FOUND. READINGS BELOW: 7: -0.2mV 10: -175.9mV 4: 167.5mV SLOPE: 175.7mV ALL PORTS TEST= FUNCTIONAL. REPLACED O-RINGS. FINAL TEST, MEETS YSI SPECIFICATIONS.					
14.1	R6150	657826		CREDITC ARD		Repair, Model 6150 6150 ROX ODO PROBE SN# 08H101103 REPORTED SYMPTOM: PREVENTATIVE MAINTENANCE TESTED AND CALIBRATED PROBE. ALL UPDATES IN PLACE REPLACED QUAD SEAL. REPLACED O-RINGS. REPLACED MEMBRANE WITH PART#606155, LOT#15B100357 TEST WIPER FUNCTION, NO DEFECT FOUND. PROBE READS: 100% IN AERATED BATH. GAIN: 1.07739 FINAL TEST, MEETS YSI SPECIFICATIONS.	9801.10.0000	EA	1	1	0
2.1	R6920V2	657826		CREDITC ARD		Repair, Model 6920V2 MODEL# 6920V2 SN# 08H100354	9801.10.0000 United States	EA	1	1	0

-For terms and conditions refer to www.ysi.com/terms_conditions

-These commodities, technology or software were exported from the United States in accordance with the Export Administration Regulations. Diversion contrary to U.S. law is prohibited.





Xylem operates under the brands of YSI, SonTek, Aanderaa, WaterLOG, and MJK

YSI Incorporated
1700/1725 Brannum Lane
Yellow Springs, OH 45387
USA
tel: 937-767-7241
fax: 937-767-9320


Packing List

Delivery Number	Packing Slip Number	Date Shipped
1393648	YSI-388164	09-MAR-15
Shipped Via: FEDEX-Air-FedEx 2nd Day		

ORDER LINE NO.	YSI ITEM NO.	ORDER NO.	CUSTOMER ITEM NO.	PO NO.	PO Line#	DESCRIPTION	HS NO./ Country Of Origin	UOM	QTY ORDERED	QTY SHIPPED	BACK ORDERED
						REPORTED SYMPTOM: P.M.					
						TEST AND CALIBRATE BOARD, NO DEFECT FOUND. TEST RTC BATTERY, NO DEFECT FOUND. FIRMWARE UP TO DATE AT 3.10. PERFORM CURRENT DRAW TEST. PASSED SLEEP, IDLE AND RUN MODE TESTING. TESTED WITH BATTERIES AND POWER SUPPLY.					
						TEST BULKHEAD FOR STRIPPED PORTS AND HIGH IMPEDANCE LEAKAGE. NO DEFECT FOUND. TEST BCR CONNECTOR AND BATTERY COMPARTMENT. NO DEFECT FOUND.					
						TEST CONDUCTIVITY/TEMPERATURE PROBE(SN#13A100461) TEMPERATURE READS 19.83°C IN A 19.8°C BATH WITH A CELL CONSTANT OF 4.91845 PROBE READS ZERO SPCOND. IN AIR. TEST PH PROBE(SN#14F) NO DEFECT FOUND. READINGS BELOW: 7: -16.9mV 10: -194.9mV 4: 153.5mV SLOPE: 178mV					
						ALL PORTS TEST=FUNCTIONAL					
						REPLACED O-RINGS.					
						FINAL TEST, MEETS YSI					

-For terms and conditions refer to www.ysi.com/terms_conditions

-These commodities, technology or software were exported from the United States in accordance with the Export Administration Regulations. Diversion contrary to U.S. law is prohibited.





Xylem operates under the brands of YSI, SonTek, Aanderaa, WaterLOG, and MJK

YSI Incorporated
1700/1725 Brannum Lane
Yellow Springs, OH 45387
USA
tel: 937-767-7241
fax: 937-767-9320

Packing List

Page 5 of 7

Delivery Number	Packing Slip Number	Date Shipped
1393648	YSI-388164	09-MAR-15
Shipped Via: FEDEX-Air-FedEx 2nd Day		

ORDER LINE NO.	YSI ITEM NO.	ORDER NO.	CUSTOMER ITEM NO.	PO NO.	PO Line#	DESCRIPTION	HS NO./ Country Of Origin	UOM	QTY ORDERED	QTY SHIPPED	BACK ORDERED
4.1	R6150	657826		CREDITC ARD		SPECIFICATIONS. Repair, Model 6150 6150 ROX ODO PROBE SN# 11E101704 REPORTED SYMPTOM: PREVENTATIVE MAINTENANCE TESTED AND CALIBRATED PROBE. ALL UPDATES IN PLACE REPLACED QUAD SEAL. REPLACED O-RINGS. REPLACED MEMBRANE WITH PART#606155, LOT#15B100357 TEST WIPER FUNCTION, NO DEFECT FOUND. PROBE READS: 100% IN AERATED BATH. GAIN: 1.10519 FINAL TEST, MEETS YSI SPECIFICATIONS.	9801.10.0000	EA	1	1	0
6.1	R6920	657826		CREDITC ARD		Repair, Model 6920 MODEL# 6920 SN# 08H100355 REPORTED SYMPTOM: P.M. TEST AND CALIBRATE BOARD, NO DEFECT FOUND. TEST RTC BATTERY, NO DEFECT FOUND. FIRMWARE UP TO DATE AT 3.10. PERFORM CURRENT DRAW TEST. PASSED SLEEP, IDLE AND RUN MODE TESTING. TESTED WITH BATTERIES AND POWER SUPPLY.	9801.10.0000	EA	1	1	0

-For terms and conditions refer to www.ysi.com/terms_conditions

-These commodities, technology or software were exported from the United States in accordance with the Export Administration Regulations. Diversion contrary to U.S. law is prohibited.





Xylem operates under the brands of YSI, SonTek, Aanderaa, WaterLOG, and MJK

YSI Incorporated
1700/1725 Brannum Lane
Yellow Springs, OH 45387
USA
tel: 937-767-7241
fax: 937-767-9320

Packing List

Page 6 of 7

Delivery Number	Packing Slip Number	Date Shipped
1393648	YSI-388164	09-MAR-15
Shipped Via: FEDEX-Air-FedEx 2nd Day		

ORDER LINE NO.	YSI ITEM NO.	ORDER NO.	CUSTOMER ITEM NO.	PO NO.	PO Line#	DESCRIPTION	HS NO./ Country Of Origin	UOM	QTY ORDERED	QTY SHIPPED	BACK ORDERED
						TEST BULKHEAD FOR STRIPPED PORTS AND HIGH IMPEDANCE LEAKAGE. NO DEFECT FOUND. TEST BCR CONNECTOR AND BATTERY COMPARTMENT. NO DEFECT FOUND.					
						TEST CONDUCTIVITY/TEMPERATURE PROBE(SN#10F100521) TEMPERATURE READS 20.44°C IN A 20.4°C BATH WITH A CELL CONSTANT OF 4.8361 PROBE READS ZERO SPCOND. IN AIR.					
						TEST PH PROBE(SN#13A) REPLACE WITH NEW DUE TO OUT OF SPEC READINGS. REPLACED AND TESTED PH PROBE(SN#15A) NO DEFECT FOUND. READINGS BELOW: 7: -2.7mV 10: -178.2mV 4: 168.3mV SLOPE: 175.5mV					
						ALL PORTS TEST=FUNCTIONAL. REPLACED O-RINGS. FINAL TEST, MEETS YSI SPECIFICATIONS.					
8.1	R6150	657826		CREDITC ARD		Repair, Model 6150 6150 ROX ODO PROBE SN# 08H101102 REPORTED SYMPTOM: PREVENTATIVE MAINTENANCE	9801.10.0000	EA	1	1	0

-For terms and conditions refer to www.ysi.com/terms_conditions

-These commodities, technology or software were exported from the United States in accordance with the Export Administration Regulations. Diversion contrary to U.S. law is prohibited.





Xylem operates under the brands of YSI, SonTek, Aanderaa, WaterLOG, and MJK

YSI Incorporated
1700/1725 Brannum Lane
Yellow Springs, OH 45387
USA
tel: 937-767-7241
fax: 937-767-9320

Packing List

Delivery Number	Packing Slip Number	Date Shipped
1393648	YSI-388164	09-MAR-15
Shipped Via: FEDEX-Air-FedEx 2nd Day		

ORDER LINE NO.	YSI ITEM NO.	ORDER NO.	CUSTOMER ITEM NO.	PO NO.	PO Line#	DESCRIPTION	HS NO./ Country Of Origin	UOM	QTY ORDERED	QTY SHIPPED	BACK ORDERED
						TESTED AND CALIBRATED PROBE. ALL UPDATES IN PLACE REPLACED QUAD SEAL. REPLACED O-RINGS. REPLACED MEMBRANE WITH PART#606155, LOT#15B100357 TEST WIPER FUNCTION, NO DEFECT FOUND. PROBE READS: 100% IN AERATED BATH. GAIN: 1.12222 FINAL TEST, MEETS YSI SPECIFICATIONS.					

GROSS WEIGHT(LB) :	18.20	NET WEIGHT (LB) :	18.20	VOLUME(CC) :	0	TOTAL CARTONS:	0
---------------------------	-------	--------------------------	-------	---------------------	---	-----------------------	---

ADDITIONAL INFO: 615558369683

ADDITIONAL NOTE: Shipping Note: Invoice Note:

Tracking Number: 

6 1 5 5 8 3 6 9 6 8 3

-For terms and conditions refer to www.ysi.com/terms_conditions

-These commodities, technology or software were exported from the United States in accordance with the Export Administration Regulations. Diversion contrary to U.S. law is prohibited.

