

**D & A INSTRUMENT**  
**40-A SETON ROAD**  
**PORT TOWNSEND, WA**  
**98368**



# CALIBRATION CERTIFICATE

CUSTOMER SERVICE (800) - 457 8352

**CUSTOMER:** Seabird Electronics  
**INSTRUMENT MODEL:** OBS-3  
**SERIAL NO:** 1591

**TYPE OF CALIBRATION:**  **FORMAZIN**  **SEDIMENT**  
**NUMBER OF STANDARD VALUES:** Two

**NOTES:**

**CALIBRATION DATA:**

STANDARD VALUE	OUTPUT
.4 NTU	0.003V +/- 0.002V
2000 NTU	4.997V +/- 0.005V

**DATE:** JULY 21, 2000

**CERTIFIED BY:**

*Dan Ewert*