



D & A Instrument Company

Sediment Instruments for All Environments

CALIBRATION CERTIFICATE

Customer: USGS-Biological Resources Division, Glacier Bay Field Station

Instrument Model: OBS-3

Serial Number: 1591

Type of Calibration: Formazin

Sediment

Number of Standard Values: Three

Calibration Data

Standard Value	Output
<u>0.2 NTU</u>	<u>0.005V +/- 0.005V</u>
<u>132 NTU</u>	<u>2.702 V +/- 0.010V</u>
<u>258 NTU</u>	<u>5.159V +/- 0.020V</u>
<u> </u>	<u> </u>
<u> </u>	<u> </u>
<u> </u>	<u> </u>

Date: July 6, 2001

Certified by: *Dan Ewert*