

CERTIFICATE OF CALIBRATION

Model Number: **LI-192SA**
UNDERWATER QUANTUM SENSOR

Serial Number: **UWQ 6011**

Calibration Constant: **4.87** (in air)
3.69 (in water)

Calibration Multiplier: **-205.34** (in air)
-271.05 (in water)

Units: microamps per 1000 $\mu\text{mol s}^{-1} \text{m}^{-2}$

Units: $\mu\text{mol s}^{-1} \text{m}^{-2}$ per microamp

Please consult the instruction manual for further information on the calibration constant and calibration multiplier. Recalibration is recommended every two years.

Date of Calibration: **February 26, 2009**

By: *Caron Deschane*

LI-COR®

Biosciences

4421 Superior Street • P.O. Box 4425 • Lincoln, Nebraska 68504 USA

Phone: 402-467-3576 • FAX: 402-467-2819

Toll-free 1-800-447-3576 (U.S. & Canada)

E-mail: envsales@licor.com

www.licor.com