

CERTIFICATE OF CALIBRATION

Model Number: **LI-192SA**
UNDERWATER QUANTUM SENSOR

Serial Number: **UWQ 6011**

Calibration Constant: **4.73** (in air)
3.58 (in water)


Calibration Multiplier: **-211.42** (in air)
-279.07 (in water)

Units: microamps per 1000 $\mu\text{mol s}^{-1} \text{m}^{-2}$

Units: $\mu\text{mol s}^{-1} \text{m}^{-2}$ per microamp

Please consult the instruction manual for further information on the calibration constant and calibration multiplier. Recalibration is recommended every two years.

Date of Calibration: **June 13, 2000**

By: 

LI-COR®

4421 Superior Street • P.O. Box 4425 • Lincoln, Nebraska 68504 USA
Phone: 402-467-3576 • FAX: 402-467-2819
Toll-free 1-800-447-3576 (U.S. & Canada)
E-mail: envsales@env.licor.com
Internet: <http://www.licor.com>