

SEA-BIRD ELECTRONICS, INC.

1808 136th Place N.E., Bellevue, Washington 98005 USA
 Phone: (425) 643 - 9866 Fax: (425) 643 - 9954 Internet: seabird@seabird.com

SENSOR SERIAL NUMBER = 775
 CALIBRATION DATE: 22-Jul-99

CONDUCTIVITY CALIBRATION DATA
 PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

GHJ COEFFICIENTS

g = -4.10533629e+00
 h = 4.90673898e-01
 i = 9.13052753e-04
 j = -1.37197723e-05
 CPcor = -9.57e-08 (nominal)
 CTcor = 3.25e-06 (nominal)

ABCDM COEFFICIENTS

a = 4.04867672e-03
 b = 4.85551546e-01
 c = -4.09881869e+00
 d = -8.50741585e-05
 m = 2.5
 CPcor = -9.57e-08 (nominal)


BATH TEMP (ITS-90 °C)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (kHz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
0.0000	0.0000	0.00000	2.88513	0.00000	0.00000
1.2390	34.5624	2.97692	8.25342	2.97684	-0.00008
4.8230	34.5624	3.29197	8.62676	3.29208	0.00011
15.5390	34.5620	4.29623	9.71951	4.29618	-0.00005
18.9130	34.5618	4.62941	10.05559	4.62944	0.00003
29.3290	34.5614	5.69991	11.06527	5.69987	-0.00004
32.7010	34.5543	6.05693	11.38187	6.05696	0.00003

$$\text{Conductivity} = (g + hf^2 + if^3 + jf^4) / [10(1 + \delta t + \epsilon p)] \text{ Siemens/meter}$$

$$\text{Conductivity} = (af^m + bf^2 + c + dt) / [10(1 + \epsilon p)] \text{ Siemens/meter}$$

t = temperature [deg C]; p = pressure [decibars]; δ = CTcor; ϵ = CPcor;

Residual = (instrument conductivity - bath conductivity) using g, h, i, j coefficients



calibration date	slope correction
⊕ 28-May-98	0.999783
⊕ 22-Jul-99	1.000000

APP NOTES

PUMP