

CTD #2



D & A Instrument Company
Sediment Instruments for All Environments

CALIBRATION CERTIFICATE

Customer: USGS - Alaska Science Center, Glacier Bay

Instrument Model: OBS-3

Serial Number: 055

Type of Calibration: Formazin Sediment

Number of Standard Values: Three

Calibration Data

Standard Value	Output
<u>0.2 NTU</u>	<u>0.005V +/- 0.005V</u>
<u>107 NTU</u>	<u>2.060V +/- 0.010V</u>
<u>250 NTU</u>	<u>4.996V +/- 0.020V</u>
<u> </u>	<u> </u>
<u> </u>	<u> </u>
<u> </u>	<u> </u>

Date: June 24, 2002

Certified by: *[Signature]*