

CERTIFICATE OF CALIBRATION

Model Number: **LI-192SA**
UNDERWATER QUANTUM SENSOR

Serial Number: **UWO 3583**

Calibration Constant: **3.82** (in air)
2.89 (in water)

LI-1000 Multiplier: **2618** (in air)
345.6 (in water)

Units: microamps per 1000 $\mu\text{mol s}^{-1} \text{m}^{-2}$

Units: $\mu\text{mol s}^{-1} \text{m}^{-2}$ per microamp

Please consult the instruction manual for further information on the calibration constant and LI-1000 Multiplier. Recalibration recommended every two years.

Date of Calibration: **JULY 29, 1996**

By: *Dave Heinicke*

 **LI-COR, inc. / LI-COR, Ltd.**
Box 4425 / Lincoln, Nebraska 68504 USA
Phone (402) 467-3576 / TWX: 910-821-8118
FAX NO. 402-467-2819