

SEA-BIRD ELECTRONICS, INC.

1808 136th Place N.E., Bellevue, Washington 98005 USA
 Phone: (425) 643 - 9866 Fax: (425) 643 - 9954 Internet: seabird@seabird.com

SENSOR SERIAL NUMBER = 775
 CALIBRATION DATE: 20-Jun-02

TEMPERATURE CALIBRATION DATA
 ITS-90 TEMPERATURE SCALE

ITS-90 COEFFICIENTS

g = 4.21290328e-03
 h = 6.03596620e-04
 i = 3.46774509e-06
 j = -2.13580847e-06
 $f_0 = 1000.000$

IPTS-68 COEFFICIENTS

a = 3.64764218e-03
 b = 5.91469265e-04
 c = 9.54540883e-06
 d = -2.13538767e-06
 $f_0 = 2571.762$

BATH TEMP (ITS-90 °C)	INSTRUMENT FREQ (Hz)	INST TEMP (ITS-90 °C)	RESIDUAL (ITS-90 °C)
0.9999	2571.762	0.9995	-0.00044
4.5000	2780.083	4.5008	0.00084
15.0001	3475.881	14.9991	-0.00105
18.5000	3732.829	18.5004	0.00036
23.9999	4162.869	24.0003	0.00037
29.0000	4582.802	29.0001	0.00015
32.5000	4893.705	32.4998	-0.00023

Temperature ITS-90 = $1/\{g + h[\ln(f_0/f)] + i[\ln^2(f_0/f)] + j[\ln^3(f_0/f)]\} - 273.15$ (°C)

Temperature IPTS-68 = $1/\{a + b[\ln(f_0/f)] + c[\ln^2(f_0/f)] + d[\ln^3(f_0/f)]\} - 273.15$ (°C)

Following the recommendation of JPOTS: T_{68} is assumed to be $1.00024 * T_{90}$ (-2 to 35 °C).

Residual = instrument temperature - bath temperature

