

# Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 6353  
CALIBRATION DATE: 07-Dec-11

SBE19plus CONDUCTIVITY CALIBRATION DATA  
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

**COEFFICIENTS:**

g = -1.011716e+000                      CPcor = -9.5700e-008  
h = 1.502430e-001                      CTcor = 3.2500e-006  
i = -1.768563e-004  
j = 3.546673e-005

BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (Hz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
22.0000	0.0000	0.00000	2596.87	0.0000	0.00000
1.0000	34.9957	2.98983	5160.27	2.9898	0.00001
4.5000	34.9749	3.29821	5354.74	3.2982	-0.00001
15.0000	34.9299	4.28407	5933.19	4.2841	-0.00001
18.5000	34.9188	4.63050	6123.25	4.6305	0.00001
24.0000	34.9053	5.19040	6418.25	5.1904	0.00001
29.0000	34.8962	5.71393	6682.01	5.7139	0.00000
32.5000	34.8890	6.08723	6863.72	6.0872	-0.00001

f = INST FREQ / 1000.0

Conductivity = (g + hf<sup>2</sup> + if<sup>3</sup> + jf<sup>4</sup>) / (1 + δt + εp) Siemens/meter

t = temperature[°C]; p = pressure[decibars]; δ = CTcor; ε = CPcor;

Residual = instrument conductivity - bath conductivity

