

Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 1685
CALIBRATION DATE: 09-Nov-16

SBE 43 OXYGEN CALIBRATION DATA

COEFFICIENTS:
Soc = 0.4658
Voffset = -0.5266
Tau20 = 1.15
A = -2.9524e-003
B = 1.6534e-004
C = -2.8866e-006
E nominal = 0.036

NOMINAL DYNAMIC COEFFICIENTS
D1 = 1.92634e-4 H1 = -3.300000e-2
D2 = -4.64803e-2 H2 = 5.00000e+3
H3 = 1.45000e+3

BATH OXYGEN (ml/l)	BATH TEMPERATURE (° C)	BATH SALINITY (PSU)	INSTRUMENT OUTPUT (volts)	INSTRUMENT OXYGEN (ml/l)	RESIDUAL (ml/l)
1.11	2.00	0.00	0.773	1.11	0.00
1.11	6.00	0.00	0.804	1.11	-0.00
1.13	12.00	0.00	0.852	1.13	-0.00
1.14	20.00	0.00	0.919	1.15	0.00
1.16	26.00	0.00	0.974	1.16	0.00
1.18	30.00	0.00	1.013	1.18	0.00
3.89	6.00	0.00	1.496	3.89	0.00
3.90	12.00	0.00	1.654	3.89	-0.00
3.90	2.00	0.00	1.396	3.90	-0.00
3.94	20.00	0.00	1.878	3.94	0.00
3.97	26.00	0.00	2.051	3.97	-0.00
3.99	30.00	0.00	2.177	3.99	0.00
6.68	2.00	0.00	2.016	6.68	-0.00
6.69	12.00	0.00	2.463	6.69	0.00
6.70	6.00	0.00	2.200	6.70	0.00
6.77	20.00	0.00	2.847	6.77	-0.00
6.80	26.00	0.00	3.141	6.81	0.00
6.82	30.00	0.00	3.346	6.82	-0.00

V = instrument output (volts); T = temperature (°C); S = salinity (PSU); K = temperature (°K)

Oxsol(T,S) = oxygen saturation (ml/l); P = pressure (dbar)

Oxygen (ml/l) = Soc * (V + Voffset) * (1.0 + A * T + B * T² + C * T³) * Oxsol(T,S) * exp(E * P / K)

Residual (ml/l) = instrument oxygen - bath oxygen

