

PO Box 518
620 Applegate St.
Philomath, OR 97370



(541) 929-5650
Fax (541) 929-5277
www.wetlabs.com

FLNTU Characterization Sheet

Date: June 17, 2009

S/N: FLNTURT-1359

Chlorophyll Scale Factor

Chlorophyll concentration expressed in $\mu\text{g/l}$ can be derived using the equation:

$$\text{CHL } (\mu\text{g/l}) = \text{Scale Factor} \times (\text{Output} - \text{Dark Counts})$$

	Analog		Digital	
Dark Counts	0.064	V	48	counts
Scale Factor (SF)	25	$\mu\text{g/lV}$	0.0301	$\mu\text{g/l/count}$
Maximum Output	4.98	V	4133	counts
Resolution	0.5	mV	1.0	counts
Ambient temperature during calibration	22.3	$^{\circ}\text{C}$		

Nephelometric Turbidity Unit (NTU) Scale Factor

Turbidity units expressed in NTU can be derived using the equation:

$$\text{NTU} = \text{Scale Factor} \times (\text{Output} - \text{Dark Counts})$$

	Analog		Digital	
Dark Counts	0.084	V	50	counts
NTU Solution Value	2.35	V	1900	counts
Scale Factor (SF)	198	NTU/V	0.2421	NTU/count
Maximum Output	4.98	V	4133	counts
Resolution	0.5	mV	1.0	counts
Ambient temperature during calibration	22.3	$^{\circ}\text{C}$		

See reverse side for definition of terms.

Dark Counts: Signal output of the meter in clean water with black tape over detector.

NTU Solution Value: Signal output of the turbidity sensor when measuring a sample of interest.

SF (CHL): Determined using the following equation: $SF = x + (\text{output} - \text{dark counts})$, where x is the concentration of the solution used during instrument characterization. SF is used to derive instrument output concentration from the raw signal output of the fluorometer.

SF (NTU): Scale factor is determined using the following equation: $SF = xx + (\text{Output} - \text{Dark counts})$, where xx is the value of a Formazin concentration. For example: $12.2 + (2011 - 50) = 0.0062$.

Maximum Output: Maximum signal output the fluorometer is capable of.

Resolution: standard deviation of 1 minute of collected data.