

Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 6353
 CALIBRATION DATE: 03-Nov-15

SBE 19plus V2 PRESSURE CALIBRATION DATA
 FSR: 870 psia S/N 2940592

COEFFICIENTS:

PA0 =	4.301058e-001	PTCA0 =	5.250179e+005
PA1 =	2.632518e-003	PTCA1 =	1.478769e+001
PA2 =	2.332431e-011	PTCA2 =	-3.226420e-001
PTEMPA0 =	-6.700965e+001	PTCB0 =	2.513975e+001
PTEMPA1 =	5.228373e+001	PTCB1 =	9.500000e-004
PTEMPA2 =	-5.213254e-001	PTCB2 =	0.000000e+000

PRESSURE SPAN CALIBRATION

PRESSURE PSIA	INST OUTPUT	THERMISTOR OUTPUT	COMPUTED PRESSURE	ERROR %FS
14.61	530582.0	1.7	14.62	0.00
179.87	593341.0	1.7	179.81	-0.01
359.98	661666.0	1.7	359.86	-0.01
539.87	729889.0	1.7	539.86	-0.00
719.88	798043.0	1.7	719.88	0.00
869.88	854755.0	1.7	869.85	-0.00
719.92	798068.0	1.7	719.95	0.00
539.93	729928.0	1.7	539.95	0.00
359.92	661692.0	1.7	359.92	0.00
179.92	593387.0	1.7	179.93	0.00
14.61	530582.0	1.7	14.62	0.00

THERMAL CORRECTION

TEMP ITS90	THERMISTOR OUTPUT	INST OUTPUT
32.50	1.94	530698.27
29.00	1.87	530708.48
24.00	1.77	530716.75
18.50	1.66	530715.47
15.00	1.59	530707.01
4.50	1.39	530621.27
1.00	1.32	530561.57

TEMP(ITS90)	SPAN(mV)
-5.00	25.14
35.00	25.17

$y = \text{thermistor output}; t = PTEMPA0 + PTEMPA1 * y + PTEMPA2 * y^2$
 $x = \text{pressure output} - PTCA0 - PTCA1 * t - PTCA2 * t^2$
 $n = x * PTCB0 / (PTCB0 + PTCB1 * t + PTCB2 * t^2)$
 $\text{pressure (psia)} = PA0 + PA1 * n + PA2 * n^2$

