

SEA-BIRD ELECTRONICS, INC.

13431 NE 20th Street, Bellevue, Washington, 98005-2010 USA

Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 6353
CALIBRATION DATE: 22-Oct-10

SBE19plus TEMPERATURE CALIBRATION DATA
ITS-90 TEMPERATURE SCALE

ITS-90 COEFFICIENTS

a0 = 1.299205e-003
a1 = 2.526166e-004
a2 = 6.731045e-007
a3 = 1.106951e-007

BATH TEMP (ITS-90)	INSTRUMENT OUTPUT(n)	INST TEMP (ITS-90)	RESIDUAL (ITS-90)
1.0000	647461.317	1.0001	0.0001
4.5000	574643.603	4.4998	-0.0002
15.0000	393513.127	15.0002	0.0002
18.5000	344868.937	18.4999	-0.0001
24.0000	278880.302	24.0000	-0.0000
29.0000	228735.333	28.9999	-0.0001
32.5000	198498.778	32.5001	0.0001

$$MV = (n - 524288) / 1.6e+007$$

$$R = (MV * 2.900e+009 + 1.024e+008) / (2.048e+004 - MV * 2.0e+005)$$

$$\text{Temperature ITS-90} = 1 / \{ a_0 + a_1[\ln(R)] + a_2[\ln^2(R)] + a_3[\ln^3(R)] \} - 273.15 \text{ (}^\circ\text{C)}$$

$$\text{Residual} = \text{instrument temperature} - \text{bath temperature}$$

